



Aesculap[®] AS Advanced Surface

7 Layers to Protect You

Aesculap[®] AS Advanced Surface

7 Layers to Protect You



Conventional monolayer coatings

showed reduced resistance against mechanical ablation, which leads to a higher risk of third body wear followed by metal ion release. $^{\left(1,2\right) }$

Multilayer coatings

This multilayer coating consisting of seven layers is unique in the market.



7-Layer Coating + Beta PE



Longevity – Ceramic Surface

65% Reduction in Wear

Wear is the number one reason for long term revision.⁽¹⁴⁾

AS knee demonstrates 65 % reduction in wear when compared to a CoCrMo prosthesis. $^{\scriptscriptstyle (3,4)}$



Wear rate (mg/Mc)





Fig.1: Wear reduction with Columbus $^{\circ}$ CR after 5 Mio cycles according ISO standard 14243-1/3 $^{(3,4,11)}$

Unmatched Hardness

Ceramic surface with a superior hardness can improve scratch resistance and implant bearing articulation.

Small scratches in CoCrMo implants are common and can lead to surface damage and higher PE wear.^(15, 16)

A hard ceramic surface improves scratch resistance.⁽¹¹⁾

The extremely hard surface shows a high resistance to scratches and also good wettability, which leads to better articulation between the polyethylene bearing surface and the femoral component. Even with the addition of cortical bone chips and bone cement particles after 5 and 5.5

million cycles (Fig. 3), no damages (scratches, nicks, etc.) could be seen on the condyle surfaces. Third body wear and the risk for mechanical ablation can be minimized this way.⁽¹¹⁾

Superior surface hardness

Hardness in GPa



No damage after extreme wear test with bone and cement particles



Fig. 3: Wear simulation under extreme conditions⁽¹¹⁾

Fig. 2: Hardness of different kind of surfaces (5-10)

Allergy Prevention – Transition Layers

20% are unsatisfied after Knee Arthroplasty Surgery.⁽¹⁷⁾ What are the Reasons for early Revision?

Patients with problems after total knee replacement have a higher level of chromium ions (p=0.001).⁽¹⁸⁾

60% of patients with poorly functionary total knees are sensitive to metal ions on the skin⁽¹⁹⁾, suggesting that the metal sensitivity is acquired through the primary arthroplasty.

Main reasons for early revision are aseptic loosening, infection and pain (Fig. 5).

How many of these could be originated in a hypersensitivity reaction?

- In a study with 1335 patients only 30% with an allergic history were detected and documented⁽²⁰⁾, which shows that metal allergies still are getting very low attention.
- Lützner et al could detect metal ions in the serum after conventional TKA.⁽²²⁾
- Metal ions may cause local and systemic toxic effects and hypersensitivity reactions.⁽²²⁾
- The metal allergy prevalence among the general population is relatively high at 13%.⁽²³⁾ The number of patients who

Main reasons for early revision < 5 years

Revision reasons in %



Fig. 5: National Joint Registry England and Wales 2010⁽²¹⁾

Allergy Prevention



Patients in need of revision are at 6 times greater risk for developing an allergic reaction⁽²⁰⁾



Fig. 6: Metal sensitivity after endoprosthesis in comparison to population⁽¹⁹⁾

Fig. 7: Metal ion reduction with AS coated implants⁽¹¹⁾

Designed for Performance

Prevention of Mechanical Ablation

A quite hard surface on the relatively soft base material (CoCrMo). Monolayer coatings showed reduced resistance against mechanical ablation.^(1,2)

The 7-layer coating is built in a way to reduce the hardness from top to bottom in a gradient way (Fig. 8). The multilayer engineering results in a very dense crystalline structure with high capacity for plastic deformation favorable to withstand corrosive environment and high stresses and strains.^(11,25)

Designed for Perfomance

Monolayer coating



Column structure with big elongated grains (low density coating)

7-layer coating: improved elastic modulus



Small grain sizes (high density coating)

Fig. 8: Hardness gradient of the 7-layer coating

Bond Coating

7 Layers to Protect You



The bonding layer between CoCrMo and transition coating forms an alloy compound with the base material promoting superior adhesion.

AS Advanced Surface is a real enhancement of coating technologies.



Beta Polyethylene Durability

Improved Age Resistance through Beta Sterilization



70% reduction in oxidation levels⁽¹³⁾

Oxidation index



ASTM F 2003: artificial aging of 10 years acc. to Kurtz et al.⁽¹²⁾: 14 days / 70°C / 5 bar O_2





Fig. 10: Oxidation level (13)



Feature	Gamma sterilization	Beta sterilization
Radiation	Lower intensity, deeper higher penetration, dosage: 2.5 Mrad – 4 Mrad	Higher intensity, concentrated, lower penetration, dosage: 2.5 Mrad – 4 Mrad
Sterilization time	Longer: 16 hours	Shorter: 15 seconds
Result	Higher content of residual free radicals leading to a higher risk of oxidation	Fewer free residual radicals after sterilization process causing less oxidation ⁽²⁷⁾
oxygenfree radicals		



Fig. 11: Gamma vs. Beta sterilisation

Enhanced Performance

Beta PE + AS Advanced Surface

Sharkey et al. showed ,Improved polyethylene or alternative bearing surfaces can certainly diminish the failure rate after knee arthroplasty.⁽⁷⁾

As known from literature, highly crosslinked polyethylenes have reduced mechanical properties in terms of elasticity and impact strength. $^{\rm (27)}$

Beta PE combines the advantage of low wear with good mechanical properties of conventional polyethylenes.



Wear rates of CR bearing offerings (ISO 14243-1/3)



Fig. 12: Wear results of CR knee systems^(11, 28-37)

The AS coating in combination with Aesculap Knee Arthroplasty systems with Beta PE yields superior wear performance.





References

- 1 Raimondi MT, Pietrabissa R. The in-vivo wear performance of prosthetic femoral heads with titanium nitride coating. Biomaterials. 2000 May;21(9):907-13.
- 2 Harman MK, Banks SA, Hodge WA. Wear analysis of a retrieved hip implant with titanium nitride coating. J Arthroplasty. 1997 Dec;12(8):938-45.
- 3 Affatato S, Spinelli M, Lopomo N, Grupp TM, Marcacci M, Toni A. Can the method of fixation influence the wear behaviour of ZrN coated unicompartmental mobile knee prostheses? Clin Biomech (Bristol, Avon). 2011 Feb;26(2):152-8. Epub 2010 Oct 8.
- 4 Grupp TM, Schwiesau T. Determination of the wear behavi or of the UNIVATION mobile knee system T018, Mar 2007.
- 5 www.medthin.com. Coating Portfolio. 30.3.2011.
- 6 TiNbN & TiN http://www.dot-coating.de/dotimplant source/beschichtung.html 30.3.2011.
- 7 Biolox Delta: Biolox Delta Nanoverbundwerkstoff für die Endoprothetik, Ceramtec 07/10.
- 8 Smith&Nephew: Oxinium: Made for Life Imagebrochure.
- 9 Zimmer PM Newsletter 11/2006.
- 10 Aesculap data on file.
- 11 Reich J, Hovy L, Lindenmaier HL, Zeller R, Schwiesau J, Thomas P, Grupp TM. Präklinische Ergebnisse beschich teter Knieimplantate für Allergiker. Orthopäde. 2010 Mai;39(5):495-502.
- 12 Kurtz SM, Muratoglu OK, Evans M, Edidin AA. Advances in the processing, sterilization, and crosslinking of ultrahigh molecular weight polyethylene for total joint arthroplasty. Biomaterials. 1999 Sep;20(18):1659-88.

- 13 Blömer W, Lohrmann E. Verschleißbeständigkeit von UHMWPE-Artikulationen in der Hüftgelenksendo prothetik. In: Weller S, Braun A, Eingartner C, Maurer F, Weise K, Winter E, Volkmann R. Das BICONTACT Hüften doprothesen-system 1987-2007. Stuttgart: Georg Thie me Verlag; 2007. p. 94-100.
- 14 Sharkey PF, Hozack WJ, Rothman RH, Shastri S, Jacoby SM. Insall Award paper. Why are total knee arthroplasties failing today? Clin Orthop Relat Res. 2002 Nov;(404):7-13.
- 15 White SE, Whiteside LA, McCarthy DS, Anthony M, Poggie RA. Simulated knee wear with cobalt chromium and oxidized zirconium knee femoral components. Clin Orthop Relat Res. 1994 Dec;(309):176-84.
- 16 Ries MD, Salehi A, Widding K, Hunter G. Polyethylene wear performance of oxidized zirconium and cobalt-chromium knee components under abrasive conditions. J Bone Joint Surg Am. 2002;84-A Suppl 2:129-35.
- 17 Bullens PH, van Loon CJ, de Waal Malefijt MC, Laan RF, Veth RP. Patient satisfaction after total knee arthroplasty: a comparison between subjective and objective outcome assessments. J Arthroplasty. 2001 Sep;16(6):740-7.
- 18 Savarino L, Tigani D, Greco M, Baldini N, Giunti A. The potential role of metal ion release as a marker of loosening in patients with total knee replacement: a cohort study. J Bone Joint Surg Br. 2010 May;92(5):634-8.
- 19 Hallab N, Merritt K, Jacobs JJ. Metal sensitivity in patients with orthopaedic implants. J Bone Joint Surg Am. 2001 Mar;83-A(3):428-36.
- 20 Rau C, Thomas P, Thomsen M. Metallallergie bei Patienten vor bzw. nach endoprothetischem Gelenkersatz. Orthopäde. 2008 Feb;37(2):102-10.
- 21 National Joint Registry of England and Wales 2010.

- 22 Luetzner J, Krummenauer F, Lengel AM, Ziegler J, Witz leb WC. Serum metal ion exposure after total knee arthroplasty. Clin Orthop Relat Res. 2007 Aug;461:136-42.
- 23 Schäfer T, Böhler E, Ruhdorfer S, Weigl L, Wessner D, Filipiak B, Wichmann HE, Ring J. Epidemiology of contact allergy in adults. Allergy. 2001 Dec;56(12):1192-6.
- 24 S. Illiger, 2021, Implant Allergy, Medical Faculty University Hospital Magdeburg, accessed 13th of October 2022, <http://www.kort.ovgu.de/en Range+of+Services/ Implant+Allergy.html.
- 25 Santana AE. Relating hardness-curve shapes with deformation mechanisms in TiAIN thin films indentation. Materials Science and Engineering A 406(2005) 11-18.
- 26 Bell CJ, Walker PS, Abeysundera MR, Simmons JM, King PM, Blunn GW. Effect of oxidation on delamination of ultrahigh-molecular-weight polyethylene tibial components. J Arthroplasty 1998 Apr;13(3):280-90.
- 27 Ries MD. Highly cross-linked polyethylene: the debate is overin opposition. J Arthroplasty. 2005 Jun;20 (4 Suppl 2):59-62.
- 28 Biomet White Paper: FDA Cleard Claim for E1 Knee Bearings - 510(k) K090528. www.biomet.com/orthopedics/ getFile.cfm?id=2657&rt=inline 22.7.2011.
- 29 Parikh A, Morrison M, Jani S. Wear testing of crosslinked and conventional UHMWPE against smooth and roughened femoral components. Orthop Res Soc, San Diego, CA, Feb 11-14, 2007, 0021.
- 30 Wang A, Yau SS, Essner A, Herrera L, Manley M, Dumbleton J. A Highly Crosslinked UHMWPE for CR and PS Total Knee arthroplasties. The Journal of Arthoplasty Vol 23 No. 4 2008.
- 31 Haider H, Alberts LR, Laurent MP, Johnson TS, Yao J, Gilbertson LN, Walker PS, Neff JR, Garvin KL. Comparison Between Force-Controlled and Displacement - Controlled

In Vitro Wear Testing on a Widely Used TKR Implant, 48th Annual Meeting of the Orthopaedic Research Society. Dallas, TX 2002, Feb.

- 32 Muratoglu OK, Bragdon, CR Jasty M, O'Connor DO, von Knoch RS, Harris WH. Knee-Simulator Testing of Conventional and Cross-Linked Polyethylene Tibial Inserts. The Jounal of Arthoplasty Vol 19 No.7 2004.
- 33 D'Lima DD , Hermida JC, Chen PC, Colwell CW. Polyethylene Wear and Variations in Knee Kinematics; Clinical Orthopaedics And Related Research; 392 (2001);124-30.
- 34 Schwiesau J. Determination of the wear behaviour of the Columbus Revision F HC Knee System Test No. T62. Tuttlingen, Jun 2008.
- 35 McEwen HMJ, Barnett PI, Bell CJ, Farrar R, Auger DD, Stone MA, Fisher J. The influence of design, materials and kinematics on the in vitro wear of total knee replacements. J Biomech, 2005;38(2):357-65.
- 36 Essner A, Herrerra L, Yau SS, Wang A, Dumbleton JH, Manley MT. Sequentially crosslinked and annealed UHMW PE Knee wear debris. 51st Orthop. Res Soc, Wahington D.C., 2005, Paper 71.
- 37 Schaerer C, Mimnaugh K, Popoola O, Seebeck J. Wear of UHMWPE tibial inserts under simulated obese patient conditions. Orthop Res Soc, New Orleans, LA, Feb 6-10, 2010, 2329.

B. Braun Australia Pty Ltd | Level 5, 7-9 Irvine Place, Bella Vista NSW 2153 Australia | Tel.1800 251 705 | info@bbraun.com | www.bbraun.com.au B. Braun New Zealand | PO Box 37353, Parnell, Auckland 1151, New Zealand | Customer Care 0800 227 286 | Fax (09) 373 5601 | www.bbraun.co.nz BAUS AE E0915 07/22 | ©2022 B. Braun Australia